



REQUEST FOR QUOTATION

RFQ No. and Date : LAB2022-08-185 / 08-09-2022
PR No. and Date : LAB2022-08-281 / 08-08-2022

The DENR - Environmental Management Bureau XI (DENR-EMBXI) Bids and Awards Committee (BAC), through its Laboratory Unit under Office of the Regional Director (ORD) will undertake a Small Value Procurement for the "Supply and Delivery of Biological Safety Cabinet" in accordance with Section 53.9 of the Implementing Rules and Regulation of the Republic Act No. 9184.

Name of Project : "Supply and Delivery of Biological Safety Cabinet"
Approved Budget of the Contract : **Six Hundred Eighty Seven Thousand Pesos Only**
(Php. 687, 000.00)
Specifications : See the attached Annex "A" for specifications
Location : DENR - Environmental Management Bureau XI, 3rd Avenue corner V. Guzman St. Brgy. 27-C, Sta. Ana, Davao City
Delivery Term : Thirty (30) days upon receipt of Purchase Order

Interested suppliers are required to submit their valid and current Mayor's Permit, PhilGEPS Registration Number, Omnibus Sworn Statement, Bank Details and Price Quotation form (Annex "A") during submission of offer/quotation.

Contract shall be awarded to the lowest calculated responsive quotation submitted, and from which complied and/or have met the minimum description stated above, and with the other terms & conditions stated in the price quotation form.

The submission of quotation and eligibility documents will be on or before **15th** of **August** at the Procurement Unit, 3rd corner V. Guzman St., Brgy. 27-C, Sta. Ana, Davao City. Open submission may be submitted, manually or through email.

Any interlineations, erasures or overwriting shall be valid only if they are signed or initialed by the bidder or his/her duly authorized representative/s.

For inquiry or for any other concerns, you may please contact us through Tel. No. (082) 234-0061, Mobile No. 09226522164, or via email embxiprocurement@emb.gov.ph / amabellecadorna@gmail.com

Very truly yours,


ANTHONY CESAR C. TAUNAN
Chief, Administrative Section

PRICE QUOTATION FORM

Date: _____

The Procurement Unit
 DENR - Environmental Management Bureau XI
 3rd Avenue corner V. Guzman St. Brgy. 27-C,
 Sta. Ana, Davao City

Sir/Madam:

After having carefully read and accepted the terms and conditions in the Request for Quotation, hereunder is our quotation/s for the item/s as follows:

| DESCRIPTION | SPECIFICATIONS | QTY/ UNIT | UNIT PRICE | TOTAL PRICE |
|--|---|--------------|---------------|----------------|
| BIOLOGICAL SAFETY CABINET, CLASS II, 4-FOOTER | <p>Protection Class: Class II, Type A2 (as per EN 12469:2000 or equivalent certification) Dimension: internal width not less than 1,220 mm; external width not greater than 1,500 mm Construction:</p> <ul style="list-style-type: none"> - With base stand and caster wheels with lock - interior and exterior must be made of high grade stainless steel preferably with antimicrobial finish - microprocessor controlled functions with digital display and alarm system for unsafe operation - tempered glass sliding sash window - side wall with covered corners - operates in 220 – 240V AC, single phase and comes with AVR of appropriate capacity <p>Work Tray: single piece stainless steel Exhaust filter: Dual ULPA filter with >99.999% efficiency Certification: must be compliant with EN 12469 or equivalent standard Inclusion: shipping, delivery, installation, testing, commissioning, and onsite end-user training Warranty: one (1) year from the date of installation Post-product service: supplier must have a service center for repair, preventive maintenance and purchase of consumables in the Philippines, preferably situated within Davao Region. Mandatory Submission: (1) complete product brochure (2) EN 12469 Certification with complete test reports of the unit being offered (3) Certification or proof of having a service center in the Philippines</p> | 1 unit | | |
| *Inclusive of Delivery/Total (Inclusive of VAT) | | | | |

(Amount in Words) _____

The above-quoted prices are inclusive of all costs and applicable taxes.

Very truly yours,

Name/ Signature of Representative_____
Name of Company_____
Contact No.

TECHNICAL SPECIFICATIONS

Bidders /suppliers must state “Comply” in the column “statement of compliance” against each of the individual parameters of each SPECIFICATION. Please do not just place check in the bidder’s “Statement of Compliance “.

| Item | Technical Specifications | Bidder’s Offer | Statement of Compliance |
|--|---|----------------|-------------------------|
| BIOLOGICAL SAFETY CABINET, CLASS II, 4-FOOTER | <p>Protection Class: Class II, Type A2 (as per EN 12469:2000 or equivalent certification)</p> <p>Dimension: internal width not less than 1,220 mm; external width not greater than 1,500 mm</p> <p>Construction:</p> <ul style="list-style-type: none"> - With base stand and caster wheels with lock - interior and exterior must be made of high grade stainless steel preferably with antimicrobial finish - microprocessor controlled functions with digital display and alarm system for unsafe operation - tempered glass sliding sash window - side wall with covered corners - operates in 220 – 240V AC, single phase and comes with AVR of appropriate capacity <p>Work Tray: single piece stainless steel</p> <p>Exhaust filter: Dual ULPA filter with >99.999% efficiency</p> <p>Certification: must be compliant with EN 12469 or equivalent standard</p> <p>Inclusion: shipping, delivery, installation, testing, commissioning, and onsite end-user training</p> <p>Warranty: one (1) year from the date of installation</p> <p>Post-product service: supplier must have a service center for repair, preventive maintenance and purchase of consumables in the Philippines, preferably situated within Davao Region.</p> <p>Mandatory Submission: (1) complete product brochure (2) EN 12469 Certification with complete test reports of the unit being offered (3) Certification or proof of having a service center in the Philippines</p> | | |

I hereby certify to comply with all the above Technical Specifications.

Name of supplier/Bidder

Signature over Printed Name of Representative

Date